

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021924\
 Data File : VW028042.D
 Acq On : 19 Feb 2024 11:52
 Operator : SY/MD
 Sample : P1469-04
 Misc : 5.05g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Feb 20 00:29:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Feb 20 00:28:09 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	277331	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	255515	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	121751	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	58041	17.869	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.480%
7) Chloroethane-d5	2.899	69	52744	21.409	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.640%
11) 1,1-Dichloroethene-d2	4.021	65	27243	19.601	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.400%
21) 2-Butanone-d5	7.069	46	24386	48.318	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.640%
24) Chloroform-d	7.648	84	139629	21.532	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.120%
26) 1,2-Dichloroethane-d4	8.307	65	66612	22.744	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.960%
32) Benzene-d6	8.270	84	285041	20.850	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.400%
36) 1,2-Dichloropropane-d6	9.270	67	76840	20.730	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	82.920%
41) Toluene-d8	10.319	98	267449	20.447	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.800%
43) trans-1,3-Dichloroprop...	10.575	79	31202	20.469	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	81.880%
47) 2-Hexanone-d5	10.922	63	21783	43.205	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.400%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	59473	21.467	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.880%
66) 1,2-Dichlorobenzene-d4	13.849	152	94271	22.781	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.120%
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	9885	2.852	ug/L	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW021924\
Data File : VW028042.D
Acq On : 19 Feb 2024 11:52
Operator : SY/MD
Sample : P1469-04
Misc : 5.05g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Feb 20 00:29:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Feb 20 00:28:09 2024
Response via : Initial Calibration





#16
Methylene chloride
Concen: 2.852 ug/L
RT: 4.923 min Scan# 49
Delta R.T. 0.000 min
Lab File: VW028042.D
Acq: 19 Feb 2024 11:52

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Tgt Ion: 84 Resp: 9885
Ion Ratio Lower Upper
84 100
86 63.5 45.4 84.4
49 81.5 66.3 123.1

