

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022222\
 Data File : VW022095.D
 Acq On : 22 Feb 2022 12:45
 Operator : SY/VA
 Sample : N1631-05
 Misc : 5.53g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBRG5

Quant Time: Feb 23 04:33:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Feb 22 00:39:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	150349	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	143536	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	71538	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	43703	24.132	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.520%
7) Chloroethane-d5	2.879	69	35673	25.586	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.360%
11) 1,1-Dichloroethene-d2	4.019	63	48428	15.135	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.560%
21) 2-Butanone-d5	7.079	46	19720	61.391	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	122.780%
24) Chloroform-d	7.646	84	78991	22.582	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.320%
26) 1,2-Dichloroethane-d4	8.305	65	45289	26.023	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.080%
32) Benzene-d6	8.274	84	155451	21.431	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.720%
36) 1,2-Dichloropropane-d6	9.274	67	43699	22.149	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.600%
41) Toluene-d8	10.323	98	149944	20.858	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.440%
43) trans-1,3-Dichloroprop...	10.579	79	19029	22.022	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.080%
47) 2-Hexanone-d5	10.926	63	17370	60.555	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.120%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	41081	25.530	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.120%
66) 1,2-Dichlorobenzene-d4	13.847	152	55968	21.687	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022222\
Data File : VW022095.D
Acq On : 22 Feb 2022 12:45
Operator : SY/VA
Sample : N1631-05
Misc : 5.53g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBRG5

Quant Time: Feb 23 04:33:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020122SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Feb 22 00:39:16 2022
Response via : Initial Calibration

