

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022322\
 Data File : VW022112.D
 Acq On : 23 Feb 2022 16:37
 Operator : SY/VA
 Sample : VW0223SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK456

Quant Time: Feb 24 03:35:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 24 03:35:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	171050	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	158600	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	77230	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	57273	25.671	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.680%
7) Chloroethane-d5	2.885	69	41673	23.740	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	94.960%
11) 1,1-Dichloroethene-d2	4.013	63	67529	19.765	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.080%
21) 2-Butanone-d5	7.080	46	17804	47.463	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.920%
24) Chloroform-d	7.647	84	94818	24.845	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.400%
26) 1,2-Dichloroethane-d4	8.305	65	49927	25.168	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.680%
32) Benzene-d6	8.275	84	191464	25.939	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.760%
36) 1,2-Dichloropropane-d6	9.274	67	52478	25.601	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	102.400%
41) Toluene-d8	10.323	98	187543	25.557	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.240%
43) trans-1,3-Dichloroprop...	10.579	79	21971	24.174	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	96.680%
47) 2-Hexanone-d5	10.926	63	15514	46.016	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.040%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	41682	23.170	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	63179	25.470	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022322\
Data File : VW022112.D
Acq On : 23 Feb 2022 16:37
Operator : SY/VA
Sample : VW0223SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK456

Quant Time: Feb 24 03:35:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 24 03:35:08 2022
Response via : Initial Calibration

