

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022322\
 Data File : VW022113.D
 Acq On : 23 Feb 2022 17:00
 Operator : SY/VA
 Sample : VW0223SBL02
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK457

Quant Time: Feb 24 03:35:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Feb 24 03:35:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	152720	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	137849	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	68289	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	42964	21.569	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.280%
7) Chloroethane-d5	2.885	69	31890	20.347	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.400%
11) 1,1-Dichloroethene-d2	4.019	63	51951	17.031	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.120%
21) 2-Butanone-d5	7.080	46	15241	45.507	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.020%
24) Chloroform-d	7.653	84	72184	21.184	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	84.720%
26) 1,2-Dichloroethane-d4	8.305	65	38448	21.707	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.840%
32) Benzene-d6	8.275	84	144497	22.523	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.080%
36) 1,2-Dichloropropane-d6	9.274	67	38766	21.759	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.040%
41) Toluene-d8	10.323	98	140261	21.991	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.960%
43) trans-1,3-Dichloroprop...	10.579	79	16533	20.929	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.720%
47) 2-Hexanone-d5	10.920	63	12660	43.204	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.400%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	32333	20.679	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.720%
66) 1,2-Dichlorobenzene-d4	13.853	152	50161	22.869	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022322\
Data File : VW022113.D
Acq On : 23 Feb 2022 17:00
Operator : SY/VA
Sample : VW0223SBL02
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK457

Quant Time: Feb 24 03:35:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Feb 24 03:35:08 2022
Response via : Initial Calibration

