

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022422\
 Data File : VW022142.D
 Acq On : 24 Feb 2022 17:57
 Operator : SY/VA
 Sample : N1631-16
 Misc : 6.03g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBRL6

Quant Time: Feb 25 04:52:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Feb 25 04:49:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	113163	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	68816	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	16378	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	31006	21.007	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.040%
7) Chloroethane-d5	2.879	69	22750	19.590	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.360%
11) 1,1-Dichloroethene-d2	4.019	63	32465	14.363	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.440%
21) 2-Butanone-d5	7.074	46	26302	105.985	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	211.980%#
24) Chloroform-d	7.647	84	66231	26.232	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	104.920%
26) 1,2-Dichloroethane-d4	8.305	65	40833	31.113	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	124.440%
32) Benzene-d6	8.275	84	109917	34.320	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	137.280%#
36) 1,2-Dichloropropane-d6	9.274	67	35200	39.576	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	158.320%#
41) Toluene-d8	10.323	98	72836	22.875	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.520%
43) trans-1,3-Dichloroprop...	10.579	79	14876	37.722	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	150.880%#
47) 2-Hexanone-d5	10.920	63	22158	151.472	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	302.940%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	36949	47.336	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	189.360%#
66) 1,2-Dichlorobenzene-d4	13.853	152	13339	25.357	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.440%
Target Compounds					Qvalue	
13) Acetone	4.129	43	2815	13.888	ug/L	98
19) 1,1-Dichloroethane	6.220	63	1383	0.665	ug/L #	90
52) Ethylbenzene	11.731	91	10993	2.638	ug/L	90
53) m,p-Xylene	11.835	106	8667	5.054	ug/L	91
54) o-Xylene	12.164	106	7178	4.449	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022422\
Data File : VW022142.D
Acq On : 24 Feb 2022 17:57
Operator : SY/VA
Sample : N1631-16
Misc : 6.03g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBRL6

Quant Time: Feb 25 04:52:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Feb 25 04:49:00 2022
Response via : Initial Calibration

