

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW022619\
 Data File : VW008844.D
 Acq On : 26 Feb 2019 21:18
 Operator : SY/VA
 Sample : K1594-40
 Misc : 6.41G/5ML/MSVOA W/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 B-9(7.5-8.5)

Quant Time: Feb 27 07:49:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	401992	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	638471	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	535618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	227363	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	251565	66.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	133.44%	
35) Dibromofluoromethane	7.88	113	198011	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
50) Toluene-d8	10.32	98	765684	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.62	95	247589	42.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.46%	
Target Compounds						
16) Acetone	4.14	43	88088	100.795	ug/l	90
17) Carbon Disulfide	4.37	76	15556	1.332	ug/l	98
20) Methylene Chloride	4.92	84	49051	7.000	ug/l	98
84) 1,2,4-Trimethylbenzene	13.26	105	15687	1.119	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW022619\
Data File : VW008844.D
Acq On : 26 Feb 2019 21:18
Operator : SY/VA
Sample : K1594-40
Misc : 6.41G/5ML/MSVOA W/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
B-9(7.5-8.5)

Quant Time: Feb 27 07:49:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

