

Data File : VW015046.D
Acq On : 26 Feb 2020 16:09
Operator : SY/VA
Sample : L1686-02
Misc : 3.76G/10ML/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB10

Quant Time: Feb 27 03:36:15 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM022420S.M
Quant Title : VOC Analysis
QLast Update : Thu Feb 27 02:01:40 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	764912	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	707129	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	258460	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	448200	23.42	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.68%
7) Chloroethane-d5	2.89	69	318830	23.12	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.48%
10) 1,1-Dichloroethene-d2	4.02	63	558157	18.75	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	75.00%
20) 2-Butanone-d5	7.07	46	149958	46.89	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.78%
24) Chloroform-d	7.65	84	626934	24.95	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.80%
26) 1,2-Dichloroethane-d4	8.30	65	352366m	25.06	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.24%
29) Benzene-d6	8.27	84	1277014	26.28	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.12%
33) 1,2-Dichloropropane-d6	9.27	67	430745	26.48	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.92%
37) Toluene-d8	10.32	98	1076823	26.57	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	106.28%
38) trans-1,3-Dichloropropene-	10.57	79	149792	24.91	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.64%
39) 2-Hexanone-d5	10.92	63	108834	50.53	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.06%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	256296	24.38	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.52%
61) 1,2-Dichlorobenzene-d4	13.85	152	274232	28.96	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	115.84%

Target Compounds

					Qvalue
5) Vinyl chloride	2.36	62	106052	5.478 ug/L	88
13) Acetone	4.12	43	11676	3.087 ug/L	83
16) Methylene chloride	4.92	84	22341	1.578 ug/L	97
18) trans-1,2-Dichloroethene	5.43	96	33676	2.849 ug/L	97
22) cis-1,2-Dichloroethene	7.16	96	2032566	168.986 ug/L	96
35) Trichloroethene	9.09	95	285384	22.428 ug/L	99
44) Toluene	10.38	91	21084	0.413 ug/L	98
47) Tetrachloroethene	10.86	164	21267	2.593 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW015046.D

Acq On : 26 Feb 2020 16:09

Operator : SY/VA

Sample : L1686-02

Misc : 3.76G/10ML/MSVOA_W/SOIL

ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB10

Quant Time: Feb 27 03:36:15 2020

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM022420S.M

Quant Title : VOC Analysis

QLast Update : Thu Feb 27 02:01:40 2020

Response via : Initial Calibration

