

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022624\
 Data File : VW028062.D
 Acq On : 26 Feb 2024 13:25
 Operator : SY/MD
 Sample : VW0226SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK435

Quant Time: Feb 27 01:32:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Feb 27 01:31:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	244147	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	233643	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	112284	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	45768	16.006	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	64.040%		
7) Chloroethane-d5	2.899	69	42772	19.721	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	78.880%		
11) 1,1-Dichloroethene-d2	4.021	65	22562	18.440	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	73.760%		
21) 2-Butanone-d5	7.075	46	27468	61.822	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	123.640%		
24) Chloroform-d	7.648	84	129911	22.757	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	91.040%		
26) 1,2-Dichloroethane-d4	8.300	65	65199	25.288	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	101.160%		
32) Benzene-d6	8.276	84	254260	20.340	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	81.360%		
36) 1,2-Dichloropropane-d6	9.270	67	71655	21.141	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	84.560%		
41) Toluene-d8	10.318	98	238167	19.913	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	79.640%		
43) trans-1,3-Dichloroprop...	10.574	79	30287	21.729	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	86.920%		
47) 2-Hexanone-d5	10.922	63	25101	54.446	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	108.900%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	64072	25.292	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	101.160%		
66) 1,2-Dichlorobenzene-d4	13.848	152	89419	23.430	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	93.720%		
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	6589	2.159	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022624\
Data File : VW028062.D
Acq On : 26 Feb 2024 13:25
Operator : SY/MD
Sample : VW0226SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK435


Quant Time: Feb 27 01:32:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Feb 27 01:31:10 2024
Response via : Initial Calibration





#16
Methylene chloride
Concen: 2.159 ug/L
RT: 4.923 min Scan# 49
Delta R.T. 0.006 min
Lab File: VW028062.D
Acq: 26 Feb 2024 13:25

Instrument :
MSVOA_W
ClientSampleId :
VBLK435



Tgt Ion: 84 Resp: 6589

Ion	Ratio	Lower	Upper
84	100		
86	57.7	45.4	84.4
49	95.6	66.3	123.1

