

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW022719\
 Data File : VW008858.D
 Acq On : 27 Feb 2019 16:49
 Operator : SY/VA
 Sample : K1594-40RE
 Misc : 5.66G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 B-9(7.5-8.5)RE

Quant Time: Feb 28 07:41:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	293066	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	443581	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	380009	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	169609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	166214	60.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.94%	
35) Dibromofluoromethane	7.88	113	136461	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
50) Toluene-d8	10.32	98	526747	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
62) 4-Bromofluorobenzene	12.62	95	171996	42.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.46%	
Target Compounds						
					Qvalue	
16) Acetone	4.13	43	56814	89.172	ug/l	99
17) Carbon Disulfide	4.38	76	12012	1.411	ug/l	99
20) Methylene Chloride	4.92	84	35114	6.771	ug/l	86
25) 2-Butanone	7.18	43	14781	16.594	ug/l	99
52) Toluene	10.39	92	8873	1.109	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW022719\
Data File : VW008858.D
Acq On : 27 Feb 2019 16:49
Operator : SY/VA
Sample : K1594-40RE
Misc : 5.66G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
B-9(7.5-8.5)RE

Quant Time: Feb 28 07:41:35 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

