

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW022720\
 Data File : VW015068.D
 Acq On : 27 Feb 2020 17:14
 Operator : SY/VA
 Sample : VIBLK13
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK13

Quant Time: Feb 28 04:43:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM022420S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Feb 28 03:59:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	696972	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	678802	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	245411	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	361916	20.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.04%
7) Chloroethane-d5	2.89	69	267511	21.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.16%
10) 1,1-Dichloroethene-d2	4.02	63	445035	16.40	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.60%
20) 2-Butanone-d5	7.07	46	180580	61.97	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.94%
24) Chloroform-d	7.65	84	542344	23.68	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.72%
26) 1,2-Dichloroethane-d4	8.31	65	337308	26.33	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.32%
29) Benzene-d6	8.27	84	1064192	22.81	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.24%
33) 1,2-Dichloropropane-d6	9.27	67	372120	23.83	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.32%
37) Toluene-d8	10.32	98	886901	22.79	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.16%
38) trans-1,3-Dichloropropene-	10.57	79	135348	23.45	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.80%
39) 2-Hexanone-d5	10.92	63	127909	61.87	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.74%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	275366	27.29	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	109.16%
61) 1,2-Dichlorobenzene-d4	13.85	152	239056	26.59	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.36%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW022720\
Data File : VW015068.D
Acq On : 27 Feb 2020 17:14
Operator : SY/VA
Sample : VIBLK13
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK13

Quant Time: Feb 28 04:43:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM022420S.M
Quant Title : VOC Analysis
QLast Update : Fri Feb 28 03:59:49 2020
Response via : Initial Calibration

