

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW022820\
 Data File : VW015081.D
 Acq On : 28 Feb 2020 15:08
 Operator : SY/VA
 Sample : L1677-16RE
 Misc : 4.79G/10ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0CD8RE

Quant Time: Feb 29 05:25:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM022420S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Feb 29 04:55:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	560140	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	469527	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	112290	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	275737	19.68	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.72%
7) Chloroethane-d5	2.89	69	211800	20.97	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.88%
10) 1,1-Dichloroethene-d2	4.02	63	349231	16.02	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.08%
20) 2-Butanone-d5	7.07	46	97468	41.62	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.24%
24) Chloroform-d	7.65	84	430462	23.39	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.56%
26) 1,2-Dichloroethane-d4	8.31	65	243428	23.64	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.56%
29) Benzene-d6	8.27	84	822242	25.48	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.92%
33) 1,2-Dichloropropane-d6	9.27	67	288969	26.75	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	107.00%
37) Toluene-d8	10.32	98	631816	23.48	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.92%
38) trans-1,3-Dichloropropene-	10.58	79	82591	20.69	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.76%
39) 2-Hexanone-d5	10.92	63	57699	40.35	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.70%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	162909	23.34	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.36%
61) 1,2-Dichlorobenzene-d4	13.85	152	111620	27.13	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	108.52%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW022820\
Data File : VW015081.D
Acq On : 28 Feb 2020 15:08
Operator : SY/VA
Sample : L1677-16RE
Misc : 4.79G/10ML/MSVOA W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0CD8RE

Quant Time: Feb 29 05:25:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM022420S.M
Quant Title : VOC Analysis
QLast Update : Sat Feb 29 04:55:00 2020
Response via : Initial Calibration

