

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022824\
 Data File : VW028079.D
 Acq On : 28 Feb 2024 16:29
 Operator : SY/MD
 Sample : P1618-05
 Misc : 5.01g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Feb 28 23:49:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Feb 28 23:27:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	254096	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	242798	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	116253	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	42170	14.170	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.680%
7) Chloroethane-d5	2.905	69	42413	18.790	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.160%
11) 1,1-Dichloroethene-d2	4.021	65	21132	16.595	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.360%
21) 2-Butanone-d5	7.081	46	28620	61.892	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.780%
24) Chloroform-d	7.648	84	122984	20.700	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.800%
26) 1,2-Dichloroethane-d4	8.307	65	62723	23.375	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.480%
32) Benzene-d6	8.276	84	238357	18.349	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	73.400%
36) 1,2-Dichloropropane-d6	9.270	67	68051	19.321	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.280%
41) Toluene-d8	10.319	98	226656	18.236	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.960%
43) trans-1,3-Dichloroprop...	10.575	79	28452	19.643	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.560%
47) 2-Hexanone-d5	10.922	63	27161	56.693	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.380%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	66298	25.184	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.720%
66) 1,2-Dichlorobenzene-d4	13.848	152	89347	22.612	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.440%
Target Compounds						
16) Methylene chloride	4.917	84	10827	3.409	ug/L	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW022824\
Data File : VW028079.D
Acq On : 28 Feb 2024 16:29
Operator : SY/MD
Sample : P1618-05
Misc : 5.01g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Feb 28 23:49:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM020524SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Feb 28 23:27:15 2024
Response via : Initial Calibration

