

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030219\
 Data File : VW008942.D
 Acq On : 02 Mar 2019 01:59
 Operator : SY/VA
 Sample : K1675-37
 Misc : 4.93G/5ML/MSVOA W/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 Z-3-14-A

Quant Time: Mar 02 05:27:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	318795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	486610	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	432648	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	198204	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	177567	59.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.78%	
35) Dibromofluoromethane	7.88	113	151743	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
50) Toluene-d8	10.32	98	583613	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
62) 4-Bromofluorobenzene	12.62	95	198852	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
Target Compounds						
16) Acetone	4.13	43	29731	42.898	ug/l	96
20) Methylene Chloride	4.92	84	34571	5.588	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW030219\
Data File : VW008942.D
Acq On : 02 Mar 2019 01:59
Operator : SY/VA
Sample : K1675-37
Misc : 4.93G/5ML/MSVOA W/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
Z-3-14-A

Quant Time: Mar 02 05:27:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

