

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030219\
 Data File : VW008946.D
 Acq On : 02 Mar 2019 03:46
 Operator : SY/VA
 Sample : K1675-09RE
 Misc : 5.20G/5ML/MSVOA W/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 Z-3-1-ARE

Quant Time: Mar 02 05:48:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	307187	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	480642	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	424616	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	191207	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	175055	60.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.52%	
35) Dibromofluoromethane	7.88	113	146153	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
50) Toluene-d8	10.32	98	572158	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
62) 4-Bromofluorobenzene	12.62	95	192157	44.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.10%	
Target Compounds						
16) Acetone	4.13	43	26705	39.988	ug/l	98
20) Methylene Chloride	4.92	84	41695	8.426	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW030219\
Data File : VW008946.D
Acq On : 02 Mar 2019 03:46
Operator : SY/VA
Sample : K1675-09RE
Misc : 5.20G/5ML/MSVOA W/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
Z-3-1-ARE

Quant Time: Mar 02 05:48:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

