

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030219\
 Data File : VW008956.D
 Acq On : 02 Mar 2019 08:13
 Operator : SY/VA
 Sample : K1670-02
 Misc : 6.26G/5ML/MSVOA W/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-GF-02-032-A

Quant Time: Mar 04 00:51:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	285642	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	444392	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	392344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	183015	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	164094	61.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.50%	
35) Dibromofluoromethane	7.88	113	141403	53.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.00%	
50) Toluene-d8	10.32	98	528364	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	12.62	95	181603	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
Target Compounds						
16) Acetone	4.13	43	27425	44.164	ug/l	95
17) Carbon Disulfide	4.37	76	19646	2.368	ug/l	99
20) Methylene Chloride	4.91	84	19543	1.425	ug/l	91
80) 1,3,5-Trimethylbenzene	12.94	105	15110	1.386	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW030219\
Data File : VW008956.D
Acq On : 02 Mar 2019 08:13
Operator : SY/VA
Sample : K1670-02
Misc : 6.26G/5ML/MSVOA W/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-GF-02-032-A

Quant Time: Mar 04 00:51:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

