

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030219\
 Data File : VW008958.D
 Acq On : 02 Mar 2019 09:07
 Operator : SY/VA
 Sample : K1671-02
 Misc : 4.88G/5ML/MSVOA W/SOIL
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-BPM-04-002-ABCD

Quant Time: Mar 04 01:00:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	58274	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	81508	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	79489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	34278	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	30642	56.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.12%	
35) Dibromofluoromethane	7.88	113	25780	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
50) Toluene-d8	10.33	98	97606	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	12.62	95	34587	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
Target Compounds						
16) Acetone	4.13	43	6079	47.984	ug/l #	85
20) Methylene Chloride	4.91	84	5279	3.731	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW030219\
Data File : VW008958.D
Acq On : 02 Mar 2019 09:07
Operator : SY/VA
Sample : K1671-02
Misc : 4.88G/5ML/MSVOA W/SOIL
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-BPM-04-002-ABCD

Quant Time: Mar 04 01:00:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

