

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030322\
 Data File : VW022159.D
 Acq On : 03 Mar 2022 13:22
 Operator : SY/VA
 Sample : N1733-04
 Misc : 6.43g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBRR8

Quant Time: Mar 04 00:32:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 04 00:31:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	196886	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	178224	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	89202	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	40131	15.627	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.520%
7) Chloroethane-d5	2.879	69	32826	16.246	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.000%
11) 1,1-Dichloroethene-d2	4.013	63	51034	12.977	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.920%
21) 2-Butanone-d5	7.074	46	23875	55.295	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.600%
24) Chloroform-d	7.647	84	98560	22.437	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.760%
26) 1,2-Dichloroethane-d4	8.305	65	52377	22.938	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.760%
32) Benzene-d6	8.275	84	181463	21.877	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.520%
36) 1,2-Dichloropropane-d6	9.275	67	53936	23.415	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.680%
41) Toluene-d8	10.323	98	175181	21.244	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.960%
43) trans-1,3-Dichloroprop...	10.579	79	21276	20.832	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.320%
47) 2-Hexanone-d5	10.927	63	19991	52.767	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.540%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	49657	24.564	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.240%
66) 1,2-Dichlorobenzene-d4	13.853	152	68050	23.751	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030322\
Data File : VW022159.D
Acq On : 03 Mar 2022 13:22
Operator : SY/VA
Sample : N1733-04
Misc : 6.43g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBRR8

Quant Time: Mar 04 00:32:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 04 00:31:34 2022
Response via : Initial Calibration

