


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030421\
Data File : VW018241.D
Acq On : 04 Mar 2021 10:07
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W030421S.M
Title : SW846 8260
Last Update : Thu Mar 04 13:49:22 2021



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.0	8503	PASS
75	95	30	60	48.1	25643	PASS
95	95	100	100	100.0	53299	PASS
96	95	5	9	6.9	3655	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	84.7	45160	PASS
175	174	5	9	7.5	3385	PASS
176	174	95	101	95.9	43315	PASS
177	176	5	9	7.5	3236	PASS