

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030519\
 Data File : VW008975.D
 Acq On : 04 Mar 2019 22:23
 Operator : SY/VA
 Sample : K1688-05
 Misc : 6.06G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-03-030119-C

Quant Time: Mar 05 07:17:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	330591	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	519927	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	465518	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	209285	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.30	65	186566	60.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.34%	
35) Dibromofluoromethane	7.88	113	157001	50.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.58%	
50) Toluene-d8	10.32	98	619755	51.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.86%	
62) 4-Bromofluorobenzene	12.62	95	213519	45.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.50%	

Target Compounds

					Qvalue	
16) Acetone	4.13	43	28674	39.897	ug/l	93
20) Methylene Chloride	4.91	84	27588	2.988	ug/l	89
30) Chloroform	7.68	83	24559	4.145	ug/l	95
69) o-Xylene	12.16	106	7778	1.185	ug/l	90
80) 1,3,5-Trimethylbenzene	12.94	105	30377	2.436	ug/l	99
84) 1,2,4-Trimethylbenzene	13.26	105	34228	2.652	ug/l	92
95) Naphthalene	15.37	128	38366	4.990	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030519\
 Data File : VW008975.D
 Acq On : 04 Mar 2019 22:23
 Operator : SY/VA
 Sample : K1688-05
 Misc : 6.06G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-03-030119-C

Quant Time: Mar 05 07:17:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

