

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030519\
 Data File : VW008976.D
 Acq On : 04 Mar 2019 22:50
 Operator : SY/VA
 Sample : K1688-07
 Misc : 6.47G/5ML/MSVOA W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-03-030119-D

Quant Time: Mar 05 07:20:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	329022	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	514638	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	452278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	194674	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	184455	59.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.54%	
35) Dibromofluoromethane	7.88	113	160253	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
50) Toluene-d8	10.32	98	609696	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	12.62	95	203191	43.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.02%	
Target Compounds						
16) Acetone	4.13	43	23126	32.331	ug/l	98
20) Methylene Chloride	4.92	84	26909	2.815	ug/l	87
69) o-Xylene	12.16	106	13247	2.078	ug/l	93
78) n-propylbenzene	12.80	91	27104	1.660	ug/l	96
80) 1,3,5-Trimethylbenzene	12.94	105	69137	5.961	ug/l	98
85) sec-Butylbenzene	13.38	105	15199	1.054	ug/l #	67
86) p-Isopropyltoluene	13.50	119	16491	1.248	ug/l	92
89) n-Butylbenzene	13.83	91	19496	1.574	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW030519\
Data File : VW008976.D
Acq On : 04 Mar 2019 22:50
Operator : SY/VA
Sample : K1688-07
Misc : 6.47G/5ML/MSVOA W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JC-03-030119-D

Quant Time: Mar 05 07:20:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

