

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW030519\
 Data File : VW008993.D
 Acq On : 05 Mar 2019 13:40
 Operator : SY/VA
 Sample : K1650-14RE
 Misc : 6.39G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 B-7(3.5-4.5)RE

Quant Time: Mar 06 08:26:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	258040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	389890	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	221105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	40086	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	156603	64.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	129.42%	
35) Dibromofluoromethane	7.88	113	136728	58.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.82%	
50) Toluene-d8	10.32	98	407722	45.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.24%	
62) 4-Bromofluorobenzene	12.62	95	64572	18.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	36.50%	
Target Compounds						
16) Acetone	4.14	43	33301	59.362	ug/l	94
30) Chloroform	7.68	83	7366	1.593	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW030519\
Data File : VW008993.D
Acq On : 05 Mar 2019 13:40
Operator : SY/VA
Sample : K1650-14RE
Misc : 6.39G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
B-7(3.5-4.5)RE

Quant Time: Mar 06 08:26:55 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

