

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030519\
 Data File : VW009003.D
 Acq On : 05 Mar 2019 18:10
 Operator : SY/VA
 Sample : K1688-07RE
 Misc : 6.43G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JC-03-030119-DRE

Quant Time: Mar 06 07:49:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 22 00:38:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	311736	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	484297	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	425804	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	181249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	176445	60.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.70%	
35) Dibromofluoromethane	7.88	113	150478	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
50) Toluene-d8	10.32	98	576433	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
62) 4-Bromofluorobenzene	12.62	95	187843	42.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.48%	
Target Compounds						
16) Acetone	4.14	43	19864	29.310	ug/l	92
69) o-Xylene	12.17	106	12359	2.059	ug/l	90
78) n-propylbenzene	12.80	91	26837	1.766	ug/l	98
80) 1,3,5-Trimethylbenzene	12.94	105	79902	7.400	ug/l	99
84) 1,2,4-Trimethylbenzene	13.26	105	11451	1.024	ug/l	95
85) sec-Butylbenzene	13.39	105	17702	1.319	ug/l #	76
86) p-Isopropyltoluene	13.50	119	16850	1.369	ug/l	94
89) n-Butylbenzene	13.82	91	20762	1.801	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW030519\
Data File : VW009003.D
Acq On : 05 Mar 2019 18:10
Operator : SY/VA
Sample : K1688-07RE
Misc : 6.43G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JC-03-030119-DRE

Quant Time: Mar 06 07:49:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W022119S.M
Quant Title : SW846 8260
QLast Update : Fri Feb 22 00:38:32 2019
Response via : Initial Calibration

