


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW030619\
Data File : VW009011.D
Acq On : 06 Mar 2019 09:52
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W030619S.M
Title : SW846 8260
Last Update : Thu Mar 07 07:22:06 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.3	46962	PASS
75	95	30	60	53.2	117261	PASS
95	95	100	100	100.0	220544	PASS
96	95	5	9	6.6	14513	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.1	189802	PASS
175	174	5	9	8.7	16478	PASS
176	174	95	101	99.3	188522	PASS
177	176	5	9	6.4	12017	PASS