

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030624\
 Data File : VW028108.D
 Acq On : 06 Mar 2024 10:25
 Operator : SY/MD
 Sample : P1677-17
 Misc : 8.94g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E0710

Quant Time: Mar 07 00:16:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 07 00:15:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	432281	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	383066	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	165055	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	97076	18.812	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.240%
7) Chloroethane-d5	2.899	69	78377	19.607	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.440%
11) 1,1-Dichloroethene-d2	4.021	65	41132	19.514	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	78.040%
21) 2-Butanone-d5	7.075	46	28180	38.283	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.560%
24) Chloroform-d	7.648	84	164868	18.007	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.040%
26) 1,2-Dichloroethane-d4	8.307	65	82039	19.253	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	77.000%
32) Benzene-d6	8.276	84	344378	19.258	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	77.040%
36) 1,2-Dichloropropane-d6	9.270	67	88974	19.257	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.040%
41) Toluene-d8	10.318	98	328186	19.474	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.880%
43) trans-1,3-Dichloroprop...	10.575	79	37618	18.580	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.320%
47) 2-Hexanone-d5	10.922	63	25935	39.852	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.700%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	65057	17.841	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	71.360%
66) 1,2-Dichlorobenzene-d4	13.848	152	84599	17.234	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	68.920%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030624\
Data File : VW028108.D
Acq On : 06 Mar 2024 10:25
Operator : SY/MD
Sample : P1677-17
Misc : 8.94g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E0710

Quant Time: Mar 07 00:16:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 07 00:15:30 2024
Response via : Initial Calibration

