

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030624\
 Data File : VW028109.D
 Acq On : 06 Mar 2024 10:49
 Operator : SY/MD
 Sample : P1677-18
 Misc : 11.48g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
ClientSampleId :
 E0711

Quant Time: Mar 07 00:16:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 07 00:15:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	418192	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	346905	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	122854	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	98878	19.807	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	79.240%		
7) Chloroethane-d5	2.899	69	82383	21.304	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	85.200%		
11) 1,1-Dichloroethene-d2	4.027	65	42710	20.946	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	83.800%		
21) 2-Butanone-d5	7.075	46	38118	53.528	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	107.060%		
24) Chloroform-d	7.648	84	176800	19.961	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	79.840%		
26) 1,2-Dichloroethane-d4	8.307	65	93058	22.575	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	90.280%		
32) Benzene-d6	8.270	84	366981	22.661	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	90.640%		
36) 1,2-Dichloropropane-d6	9.270	67	97281	23.250	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	93.000%		
41) Toluene-d8	10.319	98	339372	22.237	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	88.960%		
43) trans-1,3-Dichloroprop...	10.575	79	42376	23.111	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	92.440%		
47) 2-Hexanone-d5	10.922	63	35443	60.139	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	120.280%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	78595	23.800	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	95.200%		
66) 1,2-Dichlorobenzene-d4	13.848	152	79062	21.639	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	86.560%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030624\
 Data File : VW028109.D
 Acq On : 06 Mar 2024 10:49
 Operator : SY/MD
 Sample : P1677-18
 Misc : 11.48g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E0711

Quant Time: Mar 07 00:16:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 07 00:15:30 2024
 Response via : Initial Calibration

