

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030922\
 Data File : VW022243.D
 Acq On : 09 Mar 2022 11:08
 Operator : SY/VA
 Sample : N1818-17
 Misc : 6.15g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBSG5

Quant Time: Mar 10 04:13:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 10 04:11:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	206322	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	190152	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	89574	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	83008	30.846	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 123.400%	
7) Chloroethane-d5	2.879	69	52377	24.737	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 98.960%	
11) 1,1-Dichloroethene-d2	4.013	63	83833	20.343	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 81.360%	
21) 2-Butanone-d5	7.080	46	22705	50.181	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 100.360%	
24) Chloroform-d	7.647	84	121492	26.392	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 105.560%	
26) 1,2-Dichloroethane-d4	8.305	65	63722	26.630	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 106.520%	
32) Benzene-d6	8.275	84	245005	27.685	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 110.760%	
36) 1,2-Dichloropropane-d6	9.275	67	66712	27.145	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 108.560%	
41) Toluene-d8	10.323	98	241460	27.444	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 109.760%	
43) trans-1,3-Dichloroprop...	10.579	79	28555	26.205	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 104.800%	
47) 2-Hexanone-d5	10.921	63	19390	47.970	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 95.940%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	53160	24.647	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 98.600%	
66) 1,2-Dichlorobenzene-d4	13.847	152	80641	28.029	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 112.120%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW030922\
Data File : VW022243.D
Acq On : 09 Mar 2022 11:08
Operator : SY/VA
Sample : N1818-17
Misc : 6.15g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBSG5

Quant Time: Mar 10 04:13:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 10 04:11:04 2022
Response via : Initial Calibration

