


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW031119\
Data File : VW009076.D
Acq On : 11 Mar 2019 10:04
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W030719S.M
Title : SW846 8260
Last Update : Fri Mar 08 05:22:30 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1747

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	52520	PASS
75	95	30	60	54.0	130280	PASS
95	95	100	100	100.0	241130	PASS
96	95	5	9	6.6	16024	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	85.3	205696	PASS
175	174	5	9	8.3	17113	PASS
176	174	95	101	96.9	199381	PASS
177	176	5	9	7.0	13900	PASS