

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031122\
 Data File : VW022289.D
 Acq On : 11 Mar 2022 10:52
 Operator : SY/VA
 Sample : N1816-22
 Misc : 5.01g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Mar 11 21:02:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 11 21:01:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	141822	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	130389	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	62321	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	52130	28.181	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 112.720%	
7) Chloroethane-d5	2.879	69	36298	24.940	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 99.760%	
11) 1,1-Dichloroethene-d2	4.013	63	51635	18.228	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 72.920%	
21) 2-Butanone-d5	7.080	46	21975	70.655	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 141.320%#	
24) Chloroform-d	7.647	84	80109	25.317	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 101.280%	
26) 1,2-Dichloroethane-d4	8.305	65	44994	27.355	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 109.440%	
32) Benzene-d6	8.275	84	153208	25.247	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 101.000%	
36) 1,2-Dichloropropane-d6	9.275	67	44303	26.289	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 105.160%	
41) Toluene-d8	10.323	98	149564	24.791	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 99.160%	
43) trans-1,3-Dichloroprop...	10.579	79	18758	25.104	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 100.400%	
47) 2-Hexanone-d5	10.921	63	18057	65.147	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 130.300%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	45133	30.516	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 122.080%#	
66) 1,2-Dichlorobenzene-d4	13.853	152	58856	29.403	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 117.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031122\
Data File : VW022289.D
Acq On : 11 Mar 2022 10:52
Operator : SY/VA
Sample : N1816-22
Misc : 5.01g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Mar 11 21:02:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 11 21:01:07 2022
Response via : Initial Calibration

