

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031122\
 Data File : VW022286.D
 Acq On : 11 Mar 2022 09:40
 Operator : SY/VA
 Sample : VW0311SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK467

Quant Time: Mar 11 21:01:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 11 21:01:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	167690	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	152016	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	72270	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	56051	25.627	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.520%
7) Chloroethane-d5	2.879	69	41716	24.241	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.960%
11) 1,1-Dichloroethene-d2	4.013	63	58611	17.499	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.000%
21) 2-Butanone-d5	7.080	46	17875	48.607	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.220%
24) Chloroform-d	7.647	84	89101	23.815	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.240%
26) 1,2-Dichloroethane-d4	8.305	65	46105	23.707	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.840%
32) Benzene-d6	8.275	84	172902	24.439	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.760%
36) 1,2-Dichloropropane-d6	9.274	67	48273	24.570	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.280%
41) Toluene-d8	10.323	98	163941	23.308	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.240%
43) trans-1,3-Dichloroprop...	10.579	79	19986	22.942	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.760%
47) 2-Hexanone-d5	10.920	63	15287	47.307	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.620%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	41926	24.315	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.240%
66) 1,2-Dichlorobenzene-d4	13.853	152	61206	26.368	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031122\
Data File : VW022286.D
Acq On : 11 Mar 2022 09:40
Operator : SY/VA
Sample : VW0311SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK467

Quant Time: Mar 11 21:01:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 11 21:01:07 2022
Response via : Initial Calibration

