

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031122\
 Data File : VW022288.D
 Acq On : 11 Mar 2022 10:28
 Operator : SY/VA
 Sample : N1838-05
 Misc : 5.03g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Mar 11 21:02:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 11 21:01:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	151747	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	140349	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	68402	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	56164	28.376	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 113.520%	
7) Chloroethane-d5	2.885	69	37238	23.912	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 95.640%	
11) 1,1-Dichloroethene-d2	4.013	63	54253	17.900	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 71.600%	
21) 2-Butanone-d5	7.079	46	17508	52.611	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 105.220%	
24) Chloroform-d	7.646	84	85404	25.225	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 100.880%	
26) 1,2-Dichloroethane-d4	8.305	65	44704	25.401	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 101.600%	
32) Benzene-d6	8.274	84	162251	24.840	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 99.360%	
36) 1,2-Dichloropropane-d6	9.274	67	45781	25.238	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 100.960%	
41) Toluene-d8	10.323	98	156690	24.129	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 96.520%	
43) trans-1,3-Dichloroprop...	10.579	79	19370	24.083	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 96.320%	
47) 2-Hexanone-d5	10.926	63	14148	47.422	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 94.840%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	40078	25.175	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 100.720%	
66) 1,2-Dichlorobenzene-d4	13.847	152	58794	26.761	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 107.040%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031122\
Data File : VW022288.D
Acq On : 11 Mar 2022 10:28
Operator : SY/VA
Sample : N1838-05
Misc : 5.03g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Mar 11 21:02:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 11 21:01:07 2022
Response via : Initial Calibration

