

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031124\
 Data File : VW028144.D
 Acq On : 11 Mar 2024 11:06
 Operator : SY/MD
 Sample : VW0311SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK443

Quant Time: Mar 12 01:09:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 12 01:08:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	267097	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	240967	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	117381	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	68927	21.618	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.480%
7) Chloroethane-d5	2.905	69	53372	21.609	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.440%
11) 1,1-Dichloroethene-d2	4.027	65	27799	21.345	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	85.400%
21) 2-Butanone-d5	7.081	46	21116	46.427	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.860%
24) Chloroform-d	7.648	84	127926	22.613	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.440%
26) 1,2-Dichloroethane-d4	8.306	65	63793	24.230	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.920%
32) Benzene-d6	8.276	84	252307	22.429	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.720%
36) 1,2-Dichloropropane-d6	9.276	67	68377	23.527	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.120%
41) Toluene-d8	10.324	98	238290	22.478	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.920%
43) trans-1,3-Dichloroprop...	10.580	79	29200	22.927	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.720%
47) 2-Hexanone-d5	10.922	63	19378	47.336	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.680%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	55146	24.041	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.160%
66) 1,2-Dichlorobenzene-d4	13.848	152	87582	25.089	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031124\
Data File : VW028144.D
Acq On : 11 Mar 2024 11:06
Operator : SY/MD
Sample : VW0311SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK443

Quant Time: Mar 12 01:09:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 12 01:08:51 2024
Response via : Initial Calibration

