

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_W\DATA\VW031219\  
 Data File : VW009138.D  
 Acq On : 13 Mar 2019 06:16  
 Operator : SY/VA  
 Sample : K1871-05  
 Misc : 8.04G/5ML/MSVOA\_W/SOIL  
 ALS Vial : 30 Sample Multiplier: 1

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 TP-2

Quant Time: Mar 13 08:24:52 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_W\METHOD\82W030719S.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Mar 08 05:22:30 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 7.95  | 168  | 285144   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.84  | 114  | 450872   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.63 | 117  | 339687   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.56 | 152  | 90230    | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 33) 1,2-Dichloroethane-d4   | 8.31  | 65   | 164678   | 57.40 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 114.80% |          |
| 35) Dibromofluoromethane    | 7.88  | 113  | 139058   | 51.22 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 102.44% |          |
| 50) Toluene-d8              | 10.32 | 98   | 506823   | 45.46 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 90.92%  |          |
| 62) 4-Bromofluorobenzene    | 12.62 | 95   | 129943   | 29.61 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 59.22%  |          |


| Target Compounds               | R.T.  | QIon | Response | Conc    | Units  | Qvalue |
|--------------------------------|-------|------|----------|---------|--------|--------|
| 10) Methyl Iodide              | 4.28  | 142  | 49       | 3.202   | ug/l # | 84     |
| 11) Tert butyl alcohol         | 5.12  | 59   | 45664    | 207.715 | ug/l # | 78     |
| 16) Acetone                    | 4.14  | 43   | 53214    | 82.200  | ug/l   | 98     |
| 18) Methyl Acetate             | 4.68  | 43   | 6195     | 4.161   | ug/l   | 99     |
| 25) 2-Butanone                 | 7.17  | 43   | 20404    | 22.806  | ug/l # | 82     |
| 26) 2,2-Dichloropropane        | 7.15  | 77   | 15753    | 3.326   | ug/l # | 54     |
| 29) Tetrahydrofuran            | 7.54  | 42   | 2010     | 3.565   | ug/l # | 87     |
| 31) Cyclohexane                | 7.95  | 56   | 8405     | 1.571   | ug/l # | 38     |
| 36) 1,1-Dichloropropene        | 7.95  | 75   | 22647    | 4.804   | ug/l # | 51     |
| 37) Ethyl Acetate              | 7.17  | 43   | 20398    | 9.211   | ug/l # | 71     |
| 38) Carbon Tetrachloride       | 7.95  | 117  | 28839    | 5.923   | ug/l # | 16     |
| 49) 1,4-Dioxane                | 9.48  | 88   | 37       | 1.501   | ug/l # | 1      |
| 51) 4-Methyl-2-Pentanone       | 10.21 | 43   | 2741     | 1.309   | ug/l   | 91     |
| 55) 1,1,2-Trichloroethane      | 10.76 | 97   | 3388     | 1.452   | ug/l # | 22     |
| 59) 2-Hexanone                 | 10.94 | 43   | 5729     | 3.892   | ug/l # | 53     |
| 74) N-amyl acetate             | 12.26 | 43   | 23826    | 15.576  | ug/l # | 68     |
| 75) 1,1,2,2-Tetrachloroethane  | 12.68 | 83   | 2578     | 2.529   | ug/l # | 54     |
| 76) 1,2,3-Trichloropropane     | 12.62 | 75   | 70009    | 98.204  | ug/l # | 1      |
| 81) trans-1,4-Dichloro-2-buten | 12.62 | 75   | 70009    | 219.657 | ug/l # | 7      |
| 90) Hexachloroethane           | 14.10 | 117  | 1433     | 1.335   | ug/l # | 18     |
| 93) 1,2,4-Trichlorobenzene     | 15.14 | 180  | 6107     | 3.303   | ug/l # | 1      |
| 96) 1,2,3-Trichlorobenzene     | 15.53 | 180  | 13981    | 8.600   | ug/l # | 1      |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_W\DATA\VW031219\  
Data File : VW009138.D  
Acq On : 13 Mar 2019 06:16  
Operator : SY/VA  
Sample : K1871-05  
Misc : 8.04G/5ML/MSVOA\_W/SOIL  
ALS Vial : 30 Sample Multiplier: 1

Instrument :  
MSVOA\_W  
ClientSampled :  
TP-2

Quant Time: Mar 13 08:24:52 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_W\METHOD\82W030719S.M  
Quant Title : SW846 8260  
QLast Update : Fri Mar 08 05:22:30 2019  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.01 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Instrument :

MSVOA\_W

ClientSampleId :


TP-2

Tgt Ion:168 Resp: 285144

Ion Ratio Lower Upper

168 100

99 56.8 42.2 63.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95



150000

100000

50000

0

Time-->



#10

Methyl Iodide

Concen: 3.202 ug/l

RT: 4.28 min Scan# 389

Delta R.T. 0.01 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16


Tgt Ion:142 Resp: 49

Ion Ratio Lower Upper

142 100

127 38.8 35.8 53.6

141 0.0 11.8 17.8#



Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V

4.28

100

80


60


40

20

0

Time-->









#11

Tert butyl alcohol  
 Concen: 207.715 ug/l  
 RT: 5.12 min Scan# 528  
 Delta R.T. -0.04 min  
 Lab File: VW009138.D  
 Acq: 13 Mar 2019 06:16

Instrument :  
 MSVOA\_W  
 ClientSampleId :  
 TP-2




Tgt Ion: 59 Resp: 45664  
 Ion Ratio Lower Upper  
 59 100  
 57 2.3 8.6 12.8#




#16

Acetone  
 Concen: 82.200 ug/l  
 RT: 4.14 min Scan# 366  
 Delta R.T. 0.01 min  
 Lab File: VW009138.D  
 Acq: 13 Mar 2019 06:16

Tgt Ion: 43 Resp: 53214  
 Ion Ratio Lower Upper  
 43 100  
 58 27.1 22.5 33.7









#18  
Methyl Acetate  
Concen: 4.161 ug/l  
RT: 4.68 min Scan# 455  
Delta R.T. 0.00 min  
Lab File: VW009138.D  
Acq: 13 Mar 2019 06:16




Instrument :  
MSVOA\_W  
ClientSampled :  
TP-2


Tgt Ion: 43 Resp: 6195  
Ion Ratio Lower Upper  
43 100  
74 21.5 17.6 26.4



#25  
2-Butanone  
Concen: 22.806 ug/l  
RT: 7.17 min Scan# 863  
Delta R.T. -0.01 min  
Lab File: VW009138.D  
Acq: 13 Mar 2019 06:16

Tgt Ion: 43 Resp: 20404  
Ion Ratio Lower Upper  
43 100  
72 16.0 20.0 30.0#





#26

2,2-Dichloropropane

Concen: 3.326 ug/l

RT: 7.15 min Scan# 861

Delta R.T. -0.01 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Instrument :

MSVOA\_W

ClientSampleId :


TP-2

Tgt Ion: 77 Resp: 15753

Ion Ratio Lower Upper




77 100

97 0.0 10.9 32.9#



Abundance Ion 76.90 (76.60 to 77.60): VW

Ion 96.90 (96.60 to 97.60): VW



#29

Tetrahydrofuran

Concen: 3.565 ug/l

RT: 7.54 min Scan# 924

Delta R.T. 0.01 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16


Tgt Ion: 42 Resp: 2010

Ion Ratio Lower Upper

42 100

72 25.7 32.2 48.4#



71 37.7 30.7 46.1




Abundance Ion 42.00 (41.70 to 42.70): VW

Ion 72.00 (71.70 to 72.70): VW

Ion 71.00 (70.70 to 71.70): VW





#31

Cyclohexane

Concen: 1.571 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.01 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Instrument :

MSVOA\_W

ClientSampled :

TP-2


Tgt Ion: 56 Resp: 8405

Ion Ratio Lower Upper

56 100

69 134.1 25.5 38.3#

84 101.1 68.6 103.0






Abundance

Ion 56.00 (55.70 to 56.70): VW

Ion 69.00 (68.70 to 69.70): VW

Ion 84.00 (83.70 to 84.70): VW



#33

1,2-Dichloroethane-d4

Concen: 57.403 ug/l

RT: 8.31 min Scan# 1050

Delta R.T. 0.00 min

Lab File: VW009138.D


Acq: 13 Mar 2019 06:16

Tgt Ion: 65 Resp: 164678

Ion Ratio Lower Upper

65 100



67 51.1 0.0 103.0




Abundance

Ion 64.90 (64.60 to 65.60): VW

Ion 66.90 (66.60 to 67.60): VW









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.84 min Scan# 1138  
Delta R.T. 0.00 min  
Lab File: VW009138.D  
Acq: 13 Mar 2019 06:16

Instrument :  
MSVOA\_W  
ClientSampleId :  
TP-2

Tgt Ion:114 Resp: 450872  
Ion Ratio Lower Upper  
114 100  
63 20.7 0.0 41.2  
88 15.3 0.0 31.2




Abundance Ion 113.90 (113.60 to 114.60): V  
Ion 63.00 (62.70 to 63.70): VW  
Ion 87.90 (87.60 to 88.60): VW






#35  
Dibromofluoromethane  
Concen: 51.225 ug/l  
RT: 7.88 min Scan# 981  
Delta R.T. 0.00 min  
Lab File: VW009138.D  
Acq: 13 Mar 2019 06:16

Tgt Ion:113 Resp: 139058  
Ion Ratio Lower Upper  
113 100  
111 101.5 81.0 121.4  
192 20.7 17.1 25.7



Abundance Ion 112.80 (112.50 to 113.50): V  
Ion 110.80 (110.50 to 111.50): V  
Ion 191.70 (191.40 to 192.40): V









#36  
 1,1-Dichloropropene  
 Concen: 4.804 ug/l  
 RT: 7.95 min Scan# 991  
 Delta R.T. -0.13 min  
 Lab File: VW009138.D  
 Acq: 13 Mar 2019 06:16




Instrument :  
 MSVOA\_W  
 ClientSampled :  
 TP-2


Tgt Ion: 75 Resp: 22647  
 Ion Ratio Lower Upper  
 75 100  
 110 10.8 17.6 52.9#  
 77 0.1 25.1 37.7#



#37  
 Ethyl Acetate  
 Concen: 9.211 ug/l  
 RT: 7.17 min Scan# 863  
 Delta R.T. -0.09 min  
 Lab File: VW009138.D  
 Acq: 13 Mar 2019 06:16

Tgt Ion: 43 Resp: 20398  
 Ion Ratio Lower Upper  
 43 100  
 61 0.0 9.8 14.6#  
 70 0.0 7.2 10.8#









#38  
Carbon Tetrachloride  
Concen: 5.923 ug/l  
RT: 7.95 min Scan# 991  
Delta R.T. -0.12 min  
Lab File: VW009138.D  
Acq: 13 Mar 2019 06:16




Instrument :  
MSVOA\_W  
ClientSampled :  
TP-2


Tgt Ion: 117 Resp: 28839  
Ion Ratio Lower Upper  
117 100  
119 4.1 74.6 112.0#  
121 0.0 24.7 37.1#



#49  
1,4-Dioxane  
Concen: 1.501 ug/l  
RT: 9.48 min Scan# 1242  
Delta R.T. 0.02 min  
Lab File: VW009138.D  
Acq: 13 Mar 2019 06:16

Tgt Ion: 88 Resp: 37  
Ion Ratio Lower Upper  
88 100  
43 981.1 29.0 43.4#  
58 156.8 59.2 88.8#





#50

Toluene-d8

Concen: 45.457 ug/l

RT: 10.32 min Scan# 1381

Delta R.T. 0.00 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Instrument :

MSVOA\_W

ClientSampled :


TP-2

Tgt Ion: 98 Resp: 506823

Ion Ratio Lower Upper




98 100

100 66.1 51.9 77.9



Abundance Ion 98.00 (97.70 to 98.70): VW

Ion 100.00 (99.70 to 100.70): VW



#51

4-Methyl-2-Pentanone

Concen: 1.309 ug/l

RT: 10.21 min Scan# 1363

Delta R.T. 0.01 min

Lab File: VW009138.D


Acq: 13 Mar 2019 06:16

Tgt Ion: 43 Resp: 2741

Ion Ratio Lower Upper



43 100


58 31.0 28.8 43.2



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW





#55

1,1,2-Trichloroethane

Concen: 1.452 ug/l

RT: 10.76 min Scan# 1453

Delta R.T. -0.02 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Instrument :

MSVOA\_W

ClientSampled :

TP-2

Tgt Ion: 97 Resp: 3388


Ion Ratio Lower Upper

97 100

83 4.9 66.2 99.2#

85 9.3 43.6 65.4#

99 0.0 49.0 73.4#






Abundance Ion 96.90 (96.60 to 97.60): VW

Ion 82.90 (82.60 to 83.60): VW

Ion 84.90 (84.60 to 85.60): VW

Ion 98.90 (98.60 to 99.60): VW



#59

2-Hexanone

Concen: 3.892 ug/l

RT: 10.94 min Scan# 1482

Delta R.T. -0.03 min

Lab File: VW009138.D


Acq: 13 Mar 2019 06:16

Tgt Ion: 43 Resp: 5729

Ion Ratio Lower Upper



43 100


58 17.8 25.3 75.9#



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW









#62  
4-Bromofluorobenzene  
Concen: 29.606 ug/l  
RT: 12.62 min Scan# 1758  
Delta R.T. 0.00 min  
Lab File: VW009138.D  
Acq: 13 Mar 2019 06:16




Instrument :  
MSVOA\_W  
ClientSampled :  
TP-2


Tgt Ion: 95 Resp: 129943  
Ion Ratio Lower Upper  
95 100  
174 82.3 0.0 167.0  
176 79.4 0.0 159.4



#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.63 min Scan# 1596  
Delta R.T. 0.00 min  
Lab File: VW009138.D  
Acq: 13 Mar 2019 06:16

Tgt Ion: 117 Resp: 339687  
Ion Ratio Lower Upper  
117 100  
82 51.7 45.0 67.6  
119 31.6 26.6 39.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.56 min Scan# 1912

Delta R.T. 0.00 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Instrument :

MSVOA\_W

ClientSampled :

TP-2


Tgt Ion:152 Resp: 90230

Ion Ratio Lower Upper

152 100

115 51.9 28.9 86.7




150 142.4 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 15.576 ug/l

RT: 12.26 min Scan# 1698

Delta R.T. -0.02 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Tgt Ion: 43 Resp: 23826


Ion Ratio Lower Upper

43 100

70 24.4 30.8 46.2#

55 37.8 17.8 26.8#

61 0.2 17.4 26.2#






Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 70.00 (69.70 to 70.70): VW

Ion 55.00 (54.70 to 55.70): VW

Ion 61.00 (60.70 to 61.70): VW





#75

1,1,2,2-Tetrachloroethane

Concen: 2.529 ug/l

RT: 12.68 min Scan# 1768

Delta R.T. -0.03 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Instrument :

MSVOA\_W

ClientSampled :

TP-2





Tgt Ion: 83 Resp: 2578

Ion Ratio Lower Upper

83 100

131 0.0 5.5 16.7#

85 101.7 31.7 95.1#



#76

1,2,3-Trichloropropane

Concen: 98.204 ug/l

RT: 12.62 min Scan# 1757

Delta R.T. -0.15 min

Lab File: VW009138.D




Acq: 13 Mar 2019 06:16


Tgt Ion: 75 Resp: 70009

Ion Ratio Lower Upper

75 100

77 1.9 186.7 560.0#





#81

trans-1,4-Dichloro-2-butene

Concen: 219.657 ug/l

RT: 12.62 min Scan# 1757

Delta R.T. 0.10 min

Lab File: VW009138.D

Acq: 13 Mar 2019 06:16

Instrument :

MSVOA\_W

ClientSampled :

TP-2


Tgt Ion: 75 Resp: 70009

Ion Ratio Lower Upper

75 100

53 1.9 90.5 135.7#




89 0.9 37.5 56.3#



Abundance Ion 74.90 (74.60 to 75.60): VW

Ion 53.00 (52.70 to 53.70): VW

Ion 88.90 (88.60 to 89.60): VW



#90

Hexachloroethane

Concen: 1.335 ug/l

RT: 14.10 min Scan# 2000

Delta R.T. 0.00 min

Lab File: VW009138.D


Acq: 13 Mar 2019 06:16

Tgt Ion: 117 Resp: 1433

Ion Ratio Lower Upper



117 100


201 7.2 39.1 117.5#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V









#93  
 1,2,4-Trichlorobenzene  
 Concen: 3.303 ug/l  
 RT: 15.14 min Scan# 2171  
 Delta R.T. 0.00 min  
 Lab File: VW009138.D  
 Acq: 13 Mar 2019 06:16

Instrument :  
 MSVOA\_W  
 ClientSampled :  
 TP-2

Tgt Ion:180 Resp: 6107  
 Ion Ratio Lower Upper  
 180 100  
 182 191.7 48.1 144.3#  
 145 596.1 15.4 46.1#



#96  
 1,2,3-Trichlorobenzene  
 Concen: 8.600 ug/l  
 RT: 15.53 min Scan# 2235  
 Delta R.T. -0.03 min  
 Lab File: VW009138.D  
 Acq: 13 Mar 2019 06:16

Tgt Ion:180 Resp: 13981  
 Ion Ratio Lower Upper  
 180 100  
 182 54.7 48.1 144.3  
 145 278.5 15.8 47.3#

