

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031224\
 Data File : VW028150.D
 Acq On : 12 Mar 2024 11:01
 Operator : SY/MD
 Sample : VW0312SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK444

Quant Time: Mar 13 01:09:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Mar 13 01:09:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	272204	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	247457	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	121909	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	66376	20.428	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.720%
7) Chloroethane-d5	2.905	69	39403	15.654	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.600%
11) 1,1-Dichloroethene-d2	4.021	65	27301	20.569	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.280%
21) 2-Butanone-d5	7.075	46	19899	42.930	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.860%
24) Chloroform-d	7.642	84	123613	21.441	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.760%
26) 1,2-Dichloroethane-d4	8.301	65	60052	22.381	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	89.520%
32) Benzene-d6	8.276	84	244399	21.157	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	84.640%
36) 1,2-Dichloropropane-d6	9.270	67	65042	21.792	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.160%
41) Toluene-d8	10.319	98	234932	21.580	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.320%
43) trans-1,3-Dichloroprop...	10.575	79	28306	21.642	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.560%
47) 2-Hexanone-d5	10.922	63	18029	42.885	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.780%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	52582	22.322	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.280%
66) 1,2-Dichlorobenzene-d4	13.848	152	82482	22.750	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031224\
Data File : VW028150.D
Acq On : 12 Mar 2024 11:01
Operator : SY/MD
Sample : VW0312SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK444

Quant Time: Mar 13 01:09:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Mar 13 01:09:10 2024
Response via : Initial Calibration

