

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031324\
 Data File : VW028161.D
 Acq On : 13 Mar 2024 10:50
 Operator : SY/MD
 Sample : VW0313SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK445

Quant Time: Mar 14 01:29:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 14 01:29:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	250725	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	232351	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	114431	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	62472	20.873	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.480%
7) Chloroethane-d5	2.899	69	48619	20.970	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.880%
11) 1,1-Dichloroethene-d2	4.027	65	26159	21.397	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	85.600%
21) 2-Butanone-d5	7.075	46	24397	57.143	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	114.280%
24) Chloroform-d	7.648	84	126258	23.775	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.120%
26) 1,2-Dichloroethane-d4	8.307	65	65877	26.655	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.640%
32) Benzene-d6	8.276	84	248309	22.892	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.560%
36) 1,2-Dichloropropane-d6	9.270	67	68066	24.288	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.160%
41) Toluene-d8	10.325	98	237206	23.206	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.840%
43) trans-1,3-Dichloroprop...	10.575	79	30030	24.453	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	97.800%
47) 2-Hexanone-d5	10.922	63	22474	56.934	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.860%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	61315	27.721	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	110.880%
66) 1,2-Dichlorobenzene-d4	13.849	152	83790	24.621	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031324\
Data File : VW028161.D
Acq On : 13 Mar 2024 10:50
Operator : SY/MD
Sample : VW0313SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK445

Quant Time: Mar 14 01:29:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 14 01:29:29 2024
Response via : Initial Calibration

