

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031324\
 Data File : VW028162.D
 Acq On : 13 Mar 2024 12:49
 Operator : SY/MD
 Sample : P1742-01
 Misc : 4.97g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E0721

Quant Time: Mar 14 01:30:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 14 01:29:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	377692	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	332902	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	130341	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	87040	19.305	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.240%
7) Chloroethane-d5	2.905	69	58836	16.846	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.400%
11) 1,1-Dichloroethene-d2	4.021	65	35224	19.127	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	76.520%
21) 2-Butanone-d5	7.075	46	30648	47.653	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.300%
24) Chloroform-d	7.648	84	148729	18.592	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	74.360%
26) 1,2-Dichloroethane-d4	8.307	65	79282	21.295	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.200%
32) Benzene-d6	8.270	84	298641	19.217	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.880%
36) 1,2-Dichloropropane-d6	9.270	67	79034	19.684	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	78.720%
41) Toluene-d8	10.319	98	279242	19.067	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.280%
43) trans-1,3-Dichloroprop...	10.575	79	34853	19.808	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.240%
47) 2-Hexanone-d5	10.922	63	28353	50.133	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.260%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	70225	22.160	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.640%
66) 1,2-Dichlorobenzene-d4	13.849	152	76098	19.631	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.520%
Target Compounds						
					Qvalue	
46) Tetrachloroethene	10.861	164	57987	13.130	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031324\
Data File : VW028162.D
Acq On : 13 Mar 2024 12:49
Operator : SY/MD
Sample : P1742-01
Misc : 4.97g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E0721

Quant Time: Mar 14 01:30:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 14 01:29:29 2024
Response via : Initial Calibration

