

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031324\
 Data File : VW028169.D
 Acq On : 13 Mar 2024 15:36
 Operator : SY/MD
 Sample : P1742-07
 Misc : 5.01g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E0728

Quant Time: Mar 14 01:31:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 14 01:29:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	386794	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	349566	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	169385	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	96211	20.837	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.360%
7) Chloroethane-d5	2.899	69	54984	15.373	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.480%
11) 1,1-Dichloroethene-d2	4.027	65	40584	21.519	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	86.080%
21) 2-Butanone-d5	7.075	46	25442	38.628	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.260%
24) Chloroform-d	7.648	84	159870	19.514	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	78.040%
26) 1,2-Dichloroethane-d4	8.307	65	78729	20.649	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.600%
32) Benzene-d6	8.276	84	329233	20.175	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.720%
36) 1,2-Dichloropropane-d6	9.276	67	85569	20.295	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.200%
41) Toluene-d8	10.318	98	322129	20.947	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.800%
43) trans-1,3-Dichloroprop...	10.574	79	35534	19.232	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.920%
47) 2-Hexanone-d5	10.922	63	25253	42.523	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.040%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	68052	20.451	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.800%
66) 1,2-Dichlorobenzene-d4	13.848	152	108793	21.597	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.400%
Target Compounds						
					Qvalue	
16) Methylene chloride	4.923	84	15460	2.906	ug/L	93
46) Tetrachloroethene	10.861	164	1632	0.352	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031324\
Data File : VW028169.D
Acq On : 13 Mar 2024 15:36
Operator : SY/MD
Sample : P1742-07
Misc : 5.01g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E0728

Quant Time: Mar 14 01:31:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 14 01:29:29 2024
Response via : Initial Calibration

