

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031324\
 Data File : VW028173.D
 Acq On : 13 Mar 2024 17:11
 Operator : SY/MD
 Sample : VIBLK407
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK407

Quant Time: Mar 14 01:32:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 14 01:29:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	242494	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	219082	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	111602	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	56930	19.667	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.680%
7) Chloroethane-d5	2.905	69	46302	20.649	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.600%
11) 1,1-Dichloroethene-d2	4.027	65	23355	19.752	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.000%
21) 2-Butanone-d5	7.075	46	19112	46.284	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.560%
24) Chloroform-d	7.648	84	111880	21.783	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.120%
26) 1,2-Dichloroethane-d4	8.307	65	55953	23.408	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.640%
32) Benzene-d6	8.276	84	219787	21.490	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.960%
36) 1,2-Dichloropropane-d6	9.276	67	59707	22.596	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.400%
41) Toluene-d8	10.325	98	210652	21.856	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.440%
43) trans-1,3-Dichloroprop...	10.575	79	24509	21.166	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.680%
47) 2-Hexanone-d5	10.922	63	17631	47.370	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.740%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	50976	24.443	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.760%
66) 1,2-Dichlorobenzene-d4	13.849	152	77522	23.357	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.440%
Target Compounds						Qvalue
16) Methylene chloride	4.923	84	28863	8.653	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031324\
Data File : VW028173.D
Acq On : 13 Mar 2024 17:11
Operator : SY/MD
Sample : VIBLK407
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK407

Quant Time: Mar 14 01:32:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 14 01:29:29 2024
Response via : Initial Calibration

