

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW031419\
 Data File : VW009175.D
 Acq On : 14 Mar 2019 17:46
 Operator : SY/VA
 Sample : VW0314SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK01

Quant Time: Mar 15 07:01:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 06:53:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	502174	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	466945	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	227133	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	151652	25.27	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.08%
7) Chloroethane-d5	2.89	69	151318	21.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.20%
10) 1,1-Dichloroethene-d2	4.01	63	239210	18.44	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.76%
20) 2-Butanone-d5	7.07	46	103203	41.58	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.16%
24) Chloroform-d	7.64	84	311456	22.73	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.92%
26) 1,2-Dichloroethane-d4	8.30	65	186369	22.57	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.28%
29) Benzene-d6	8.27	84	621707	24.14	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.56%
33) 1,2-Dichloropropane-d6	9.27	67	182250	23.56	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	94.24%
37) Toluene-d8	10.32	98	579265	24.33	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.32%
38) trans-1,3-Dichloropropene-	10.57	79	86572	22.29	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.16%
39) 2-Hexanone-d5	10.93	63	78398	41.85	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.70%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	160703	21.10	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.40%
61) 1,2-Dichlorobenzene-d4	13.86	152	213639	24.53	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.12%

Target Compounds					Ovalue
13) Acetone	4.12	43	4506	2.214 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW031419\
Data File : VW009175.D
Acq On : 14 Mar 2019 17:46
Operator : SY/VA
Sample : VW0314SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK01

Quant Time: Mar 15 07:01:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 06:53:55 2019
Response via : Initial Calibration

