

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031422\
 Data File : VW022313.D
 Acq On : 14 Mar 2022 12:04
 Operator : SY/VA
 Sample : N1818-25
 Misc : 5.01g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Mar 15 01:04:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 15 01:03:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	134200	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	129571	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	63116	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	37462	21.402	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.600%
7) Chloroethane-d5	2.879	69	23147	16.807	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.240%
11) 1,1-Dichloroethene-d2	4.013	63	35571	13.270	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.080%
21) 2-Butanone-d5	7.080	46	15598	53.000	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.000%
24) Chloroform-d	7.647	84	64520	21.548	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	8.305	65	35163	22.592	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.360%
32) Benzene-d6	8.275	84	120408	19.967	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.880%
36) 1,2-Dichloropropane-d6	9.268	67	35671	21.301	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.200%
41) Toluene-d8	10.323	98	113513	18.934	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.720%
43) trans-1,3-Dichloroprop...	10.579	79	14648	19.727	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.920%
47) 2-Hexanone-d5	10.920	63	12147	44.101	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.200%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	34606	23.546	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	45286	22.339	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031422\
Data File : VW022313.D
Acq On : 14 Mar 2022 12:04
Operator : SY/VA
Sample : N1818-25
Misc : 5.01g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Mar 15 01:04:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 15 01:03:29 2022
Response via : Initial Calibration

