

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031422\
 Data File : VW022326.D
 Acq On : 14 Mar 2022 17:37
 Operator : SY/VA
 Sample : N1883-01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBRZ9

Quant Time: Mar 15 01:07:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 15 01:03:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	156201	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	148440	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	71807	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	44677	21.929	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.720%
7) Chloroethane-d5	2.873	69	34151	21.304	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.200%
11) 1,1-Dichloroethene-d2	4.007	63	40530	12.991	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.960%
21) 2-Butanone-d5	7.074	46	23488	68.568	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	137.140%#
24) Chloroform-d	7.647	84	74046	21.247	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.000%
26) 1,2-Dichloroethane-d4	8.305	65	42738	23.592	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.360%
32) Benzene-d6	8.275	84	136641	19.779	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.120%
36) 1,2-Dichloropropane-d6	9.274	67	42101	21.944	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.760%
41) Toluene-d8	10.323	98	124297	18.098	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.400%
43) trans-1,3-Dichloroprop...	10.579	79	18244	21.447	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.800%
47) 2-Hexanone-d5	10.920	63	19244	60.987	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.980%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	45823	27.215	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	108.880%
66) 1,2-Dichlorobenzene-d4	13.853	152	51554	22.353	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031422\
Data File : VW022326.D
Acq On : 14 Mar 2022 17:37
Operator : SY/VA
Sample : N1883-01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBRZ9

Quant Time: Mar 15 01:07:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 15 01:03:29 2022
Response via : Initial Calibration

