

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031422\
 Data File : VW022327.D
 Acq On : 14 Mar 2022 18:01
 Operator : SY/VA
 Sample : N1883-02
 Misc : 5.42g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBSA1

Quant Time: Mar 15 01:07:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 15 01:03:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	165725	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	155751	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	75889	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	43775	20.252	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.000%
7) Chloroethane-d5	2.873	69	35216	20.706	ug/L	-0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.840%
11) 1,1-Dichloroethene-d2	4.007	63	41606	12.569	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	50.280%
21) 2-Butanone-d5	7.074	46	24512	67.445	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	134.900%
24) Chloroform-d	7.647	84	80375	21.737	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.960%
26) 1,2-Dichloroethane-d4	8.299	65	46203	24.039	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.160%
32) Benzene-d6	8.269	84	144664	19.957	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.840%
36) 1,2-Dichloropropane-d6	9.275	67	44740	22.225	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.920%
41) Toluene-d8	10.323	98	136001	18.872	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.480%
43) trans-1,3-Dichloroprop...	10.573	79	18874	21.146	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.600%
47) 2-Hexanone-d5	10.921	63	19932	60.202	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	120.400%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	49000	27.736	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	110.960%
66) 1,2-Dichlorobenzene-d4	13.847	152	57083	23.419	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031422\
Data File : VW022327.D
Acq On : 14 Mar 2022 18:01
Operator : SY/VA
Sample : N1883-02
Misc : 5.42g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBSA1

Quant Time: Mar 15 01:07:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 15 01:03:29 2022
Response via : Initial Calibration

