

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031422\
 Data File : VW022315.D
 Acq On : 14 Mar 2022 13:11
 Operator : SY/VA
 Sample : N1819-23
 Misc : 5.02g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Mar 15 01:05:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 15 01:03:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	142896	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	129617	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	63273	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	42785	22.956	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.840%
7) Chloroethane-d5	2.879	69	28978	19.761	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.040%
11) 1,1-Dichloroethene-d2	4.013	63	42123	14.758	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.040%
21) 2-Butanone-d5	7.080	46	17276	55.129	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	110.260%
24) Chloroform-d	7.647	84	72956	22.883	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.520%
26) 1,2-Dichloroethane-d4	8.305	65	39627	23.911	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.640%
32) Benzene-d6	8.275	84	137482	22.791	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.160%
36) 1,2-Dichloropropane-d6	9.274	67	40098	23.936	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.760%
41) Toluene-d8	10.323	98	126151	21.035	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.120%
43) trans-1,3-Dichloroprop...	10.579	79	16663	22.433	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.720%
47) 2-Hexanone-d5	10.920	63	14484	52.567	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.140%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	39211	26.670	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	106.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	51288	25.237	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031422\
Data File : VW022315.D
Acq On : 14 Mar 2022 13:11
Operator : SY/VA
Sample : N1819-23
Misc : 5.02g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Mar 15 01:05:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 15 01:03:29 2022
Response via : Initial Calibration

