

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031622\
 Data File : VW022354.D
 Acq On : 16 Mar 2022 13:19
 Operator : SY/VA
 Sample : VW0316SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK470

Quant Time: Mar 17 00:19:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 17 00:17:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	130913	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	125750	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	63951	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	36360	21.294	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.160%
7) Chloroethane-d5	2.885	69	27863	20.739	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.960%
11) 1,1-Dichloroethene-d2	4.013	63	39310	15.033	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.120%
21) 2-Butanone-d5	7.080	46	17807m	62.025	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	124.060%
24) Chloroform-d	7.647	84	71930	24.626	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.520%
26) 1,2-Dichloroethane-d4	8.305	65	37635	24.788	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.160%
32) Benzene-d6	8.275	84	131542	22.477	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.920%
36) 1,2-Dichloropropane-d6	9.275	67	39017	24.006	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.040%
41) Toluene-d8	10.323	98	126161	21.683	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.720%
43) trans-1,3-Dichloroprop...	10.579	79	16344	22.680	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.720%
47) 2-Hexanone-d5	10.921	63	14537	54.382	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.760%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	39732	27.855	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.440%
66) 1,2-Dichlorobenzene-d4	13.853	152	53070	25.837	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031622\
Data File : VW022354.D
Acq On : 16 Mar 2022 13:19
Operator : SY/VA
Sample : VW0316SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK470

Quant Time: Mar 17 00:19:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 17 00:17:26 2022
Response via : Initial Calibration

