

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031622\
 Data File : VW022357.D
 Acq On : 16 Mar 2022 15:13
 Operator : SY/VA
 Sample : VIBLK407
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK407

Quant Time: Mar 17 00:19:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 17 00:17:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	156587	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	142235	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	69202	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	39236	19.211	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.840%
7) Chloroethane-d5	2.885	69	30425	18.933	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.720%
11) 1,1-Dichloroethene-d2	4.013	63	40566	12.970	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	51.880%
21) 2-Butanone-d5	7.086	46	14529	42.310	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.620%
24) Chloroform-d	7.646	84	72017	20.614	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.440%
26) 1,2-Dichloroethane-d4	8.305	65	36787	20.257	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.040%
32) Benzene-d6	8.274	84	133033	20.097	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	80.400%
36) 1,2-Dichloropropane-d6	9.274	67	40204	21.870	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.480%
41) Toluene-d8	10.323	98	123938	18.832	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	75.320%
43) trans-1,3-Dichloroprop...	10.579	79	15702	19.264	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.040%
47) 2-Hexanone-d5	10.920	63	11258	37.234	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.460%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	35243	21.845	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.360%
66) 1,2-Dichlorobenzene-d4	13.853	152	49860	22.432	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031622\
Data File : VW022357.D
Acq On : 16 Mar 2022 15:13
Operator : SY/VA
Sample : VIBLK407
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK407

Quant Time: Mar 17 00:19:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM022322SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 17 00:17:26 2022
Response via : Initial Calibration

