

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW031919\
 Data File : VW009267.D
 Acq On : 20 Mar 2019 05:06
 Operator : SY/VA
 Sample : K1984-04
 Misc : 5.01G/5ML/MSVOA W/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-11(14-15)

Quant Time: Mar 20 08:58:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W031919S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 20 07:22:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	293015	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	484769	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	444449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	211655	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	189823	60.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.80%	
35) Dibromofluoromethane	7.88	113	160993	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
50) Toluene-d8	10.32	98	602610	53.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.84%	
62) 4-Bromofluorobenzene	12.62	95	213885	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
Target Compounds						
16) Acetone	4.12	43	12873	20.132	ug/l #	88
17) Carbon Disulfide	4.38	76	113971	14.619	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW031919\
Data File : VW009267.D
Acq On : 20 Mar 2019 05:06
Operator : SY/VA
Sample : K1984-04
Misc : 5.01G/5ML/MSVOA W/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-11(14-15)

Quant Time: Mar 20 08:58:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W031919S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 20 07:22:50 2019
Response via : Initial Calibration

