

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031921\
 Data File : VW018404.D
 Acq On : 19 Mar 2021 22:27
 Operator : SY/VA
 Sample : M1698-17
 Misc : 5.53G/10ML/MSVOA_W/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG4L5

Quant Time: Mar 20 04:00:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 20 03:58:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	354939	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	248707	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.560	152	60867	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	143824	29.12	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	116.48%
7) Chloroethane-d5	2.891	69	130093	30.09	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	120.36%
10) 1,1-Dichloroethene-d2	4.019	63	153558	18.01	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	72.04%
20) 2-Butanone-d5	7.080	46	74956	68.16	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	136.32%#
24) Chloroform-d	7.647	84	222652	24.63	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.52%
26) 1,2-Dichloroethane-d4	8.305	65	129979	24.63	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.52%
29) Benzene-d6	8.274	84	427422	32.33	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	129.32%
33) 1,2-Dichloropropane-d6	9.274	67	126937	33.28	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	133.12%#
37) Toluene-d8	10.323	98	344266	27.04	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	108.16%
38) trans-1,3-Dichloroprop...	10.579	79	30596	15.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	62.24%
39) 2-Hexanone-d5	10.926	63	52968	70.97	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	141.94%#
48) 1,1,2,2-Tetrachloroeth...	12.688	84	97504	28.58	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	114.32%
61) 1,2-Dichlorobenzene-d4	13.853	152	57331	26.49	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	105.96%
Target Compounds						
					Qvalue	
13) Acetone	4.123	43	22963	29.07	ug/L	97
14) Carbon disulfide	4.385	76	3277	0.24	ug/L #	92
16) Methylene chloride	4.915	84	24679	4.24	ug/L	95
21) 2-Butanone	7.177	43	5435	4.33	ug/L	90
47) Tetrachloroethene	10.859	164	18673	6.18	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW031921\
Data File : VW018404.D
Acq On : 19 Mar 2021 22:27
Operator : SY/VA
Sample : M1698-17
Misc : 5.53G/10ML/MSVOA_W/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG4L5

Quant Time: Mar 20 04:00:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 20 03:58:53 2021
Response via : Initial Calibration

