

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW032219\
 Data File : VW009362.D
 Acq On : 22 Mar 2019 23:49
 Operator : SY/VA
 Sample : K2031-02
 Misc : 5.63G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF5S2

Manual Integrations
 APPROVED

MMDadoda
 3/25/2019 8:29:41 AM

Quant Time: Mar 23 02:01:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 23 01:25:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	405908	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	369912	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	152338	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	107364	22.13	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.52%
7) Chloroethane-d5	2.89	69	109724	19.33	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.32%
10) 1,1-Dichloroethene-d2	4.01	63	165723m	15.81	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.24%
20) 2-Butanone-d5	7.07	46	84537	42.14	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.28%
24) Chloroform-d	7.64	84	257330	23.23	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.92%
26) 1,2-Dichloroethane-d4	8.31	65	153433	22.99	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.96%
29) Benzene-d6	8.27	84	464726	22.78	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.12%
33) 1,2-Dichloropropane-d6	9.27	67	150690	24.59	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.36%
37) Toluene-d8	10.32	98	403732	21.40	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.60%
38) trans-1,3-Dichloropropene-	10.58	79	59612	19.37	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.48%
39) 2-Hexanone-d5	10.93	63	60933	41.06	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	82.12%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	129802	21.51	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.04%
61) 1,2-Dichlorobenzene-d4	13.86	152	136397	23.35	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.40%

Target Compounds					Ovalue
13) Acetone	4.12	43	36575	22.233 ug/L	100
14) Carbon disulfide	4.37	76	16297	1.211 ug/L #	94
16) Methylene chloride	4.91	84	18789m	3.174 ug/L	
21) 2-Butanone	7.17	43	11143	4.906 ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW032219\
Data File : VW009362.D
Acq On : 22 Mar 2019 23:49
Operator : SY/VA
Sample : K2031-02
Misc : 5.63G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF5S2

Manual Integrations
APPROVED

MMDadoda
3/25/2019 8:29:41 AM

Quant Time: Mar 23 02:01:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 23 01:25:26 2019
Response via : Initial Calibration

