


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032321\
Data File : VW018431.D
Acq On : 23 Mar 2021 17:04
Operator : SY/VA
Sample : M1781-10
Misc : 5.01G/10ML/MSVOA_W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BG4R9

Manual Integrations
APPROVED

SAM
3/26/2021 6:06:22 PM

Quant Time: Mar 24 01:27:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 24 01:23:39 2021
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032321\
 Data File : VW018431.D
 Acq On : 23 Mar 2021 17:04
 Operator : SY/VA
 Sample : M1781-10
 Misc : 5.01G/10ML/MSVOA_W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampled :
 BG4R9

Manual Integrations
 APPROVED

SAM
 3/26/2021 6:06:22 PM

Quant Time: Mar 24 01:27:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 24 01:23:39 2021
 Response via : Initial Calibration



TIC: VW018431.D\data.ms

(16) Methylene chloride (T)

4.922min (0.000) 1.03 ug/L m

response 8678

Ion	Exp%	Act%
84.00	100.00	100.00
49.00	106.30	104.50
86.00	65.90	70.51
0.00	0.00	0.00

> MB
 3/30/21

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032321\
 Data File : VW018431.D
 Acq On : 23 Mar 2021 17:04
 Operator : SY/VA
 Sample : M1781-10
 Misc : 5.01G/10ML/MSVOA_W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BG4R9

Manual Integrations
 APPROVED

SAM
 3/26/2021 6:06:22 PM

Quant Time: Mar 24 01:27:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 24 01:23:39 2021
 Response via : Initial Calibration



TIC: VW018431.D\data.ms

(16) Methylene chloride (T)

4.922min (0.000) 0.68 ug/L

response 5713

Ion	Exp%	Act%
84.00	100.00	100.00
49.00	106.30	104.50
86.00	65.90	70.51
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032321\
 Data File : VW018431.D
 Acq On : 23 Mar 2021 17:04
 Operator : SY/VA
 Sample : M1781-10
 Misc : 5.01G/10ML/MSVOA_W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampled :
 BG4R9

Manual Integrations
 APPROVED

SAM
 3/26/2021 6:06:22 PM

Quant Time: Mar 24 01:27:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 24 01:23:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	514618	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	428128	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.554	152	181685	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	156129	21.80	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	87.20%		
7) Chloroethane-d5	2.898	69	147732	23.57	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	94.28%		
10) 1,1-Dichloroethene-d2	4.019	63	199382	16.12	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	64.48%		
20) 2-Butanone-d5	7.080	46	61614	38.65	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	77.30%		
24) Chloroform-d	7.647	84	264079	20.15	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	80.60%		
26) 1,2-Dichloroethane-d4	8.305	65	154594	20.20	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	80.80%		
29) Benzene-d6	8.275	84	515676	22.66	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	90.64%		
33) 1,2-Dichloropropane-d6	9.275	67	147042	22.40	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	89.60%		
37) Toluene-d8	10.323	98	475932	21.72	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	86.88%		
38) trans-1,3-Dichloroprop...	10.579	79	61070	18.04	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	72.16%		
39) 2-Hexanone-d5	10.927	63	50629	39.41	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	78.82%		
48) 1,1,2,2-Tetrachloroeth...	12.695	84	115639	19.69	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	78.76%		
61) 1,2-Dichlorobenzene-d4	13.853	152	133321	20.64	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	82.56%		

Target Compounds					Qvalue
13) Acetone	4.154	43	6859	5.99	ug/L 90
14) Carbon disulfide	4.391	76	14926	0.77	ug/L 96
16) Methylene chloride	4.922	84	8678m	1.03	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

7 m0
 3/30/21