

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032421\
 Data File : VW018437.D
 Acq On : 24 Mar 2021 10:39
 Operator : SY/VA
 Sample : VW0324SBL01
 Misc : 5.00G/10ML/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK21

Manual Integrations
 APPROVED

MMDadoda
 3/26/2021 9:31:12 AM

Quant Time: Mar 25 01:24:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 25 01:23:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	477764	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	440896	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.554	152	210815	25.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	142504	21.43	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	85.72%		
7) Chloroethane-d5	2.891	69	134137	23.05	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	92.20%		
10) 1,1-Dichloroethene-d2	4.019	63	179543	15.64	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	62.56%		
20) 2-Butanone-d5	7.074	46	72214	48.79	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	97.58%		
24) Chloroform-d	7.647	84	272020	22.35	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	89.40%		
26) 1,2-Dichloroethane-d4	8.305	65	167537	23.58	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	94.32%		
29) Benzene-d6	8.275	84	529893	22.61	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	90.44%		
33) 1,2-Dichloropropane-d6	9.274	67	152903	22.62	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	90.48%		
37) Toluene-d8	10.323	98	500588	22.18	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	88.72%		
38) trans-1,3-Dichloroprop...	10.579	79	75587	21.68	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	86.72%		
39) 2-Hexanone-d5	10.927	63	59977	45.33	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	90.66%		
48) 1,1,2,2-Tetrachloroeth...	12.688	84	139811	23.12	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	92.48%		
61) 1,2-Dichlorobenzene-d4	13.853	152	169822	22.66	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	90.64%		
Target Compounds						Qvalue
16) Methylene chloride	4.915	84	6679m	0.85	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032421\
Data File : VW018437.D
Acq On : 24 Mar 2021 10:39
Operator : SY/VA
Sample : VW0324SBL01
Misc : 5.00G/10ML/MSVOA_W/SOIL
ALS Vial : 3 Sample Multiplier: 1


Instrument :
MSVOA_W
ClientSampleId :
VBLK21

Manual Integrations
APPROVED

MMDadoda
3/26/2021 9:31:12 AM

Quant Time: Mar 25 01:24:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SOM2WLM031921S.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 25 01:23:35 2021
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.85 ug/L m
RT: 4.915 min Scan# 494
Delta R.T. 0.006 min
Lab File: VW018437.D
Acq: 24 Mar 2021 10:39

Tgt Ion: 84 Resp: 6679

Ion	Ratio	Lower	Upper
84	100		
49	101.2	74.4	138.2
86	47.3	46.1	85.7

Instrument :
MSVOA_W
ClientSampled :
VBLK21

Manual Integrations
APPROVED

MMDadoda
3/26/2021 9:31:12 AM

