

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW032519\
 Data File : VW009446.D
 Acq On : 25 Mar 2019 23:57
 Operator : SY/VA
 Sample : K2044-14
 Misc : 5.38G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BF5W3

Quant Time: Apr 08 10:48:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 26 09:00:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	376462	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	292724	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	79423	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	84822	18.85	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.40%
7) Chloroethane-d5	2.89	69	97701	18.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.24%
10) 1,1-Dichloroethene-d2	4.01	63	139196	14.32	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.28%
20) 2-Butanone-d5	7.08	46	111170	59.75	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	119.50%
24) Chloroform-d	7.64	84	236873	23.06	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.24%
26) 1,2-Dichloroethane-d4	8.31	65	157613	25.47	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.88%
29) Benzene-d6	8.27	84	410351	25.42	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.68%
33) 1,2-Dichloropropane-d6	9.27	67	135488	27.94	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	111.76%
37) Toluene-d8	10.32	98	313116	20.98	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.92%
38) trans-1,3-Dichloropropene-	10.57	79	58086	23.85	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	95.40%
39) 2-Hexanone-d5	10.93	63	79519	67.71	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	135.42%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	112772	23.62	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.48%
61) 1,2-Dichlorobenzene-d4	13.86	152	64348	21.13	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.52%

Target Compounds					Ovalue
13) Acetone	4.12	43	30319	19.872 ug/L	97
14) Carbon disulfide	4.37	76	11094	0.889 ug/L	99
16) Methylene chloride	4.92	84	23913	4.356 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW032519\
Data File : VW009446.D
Acq On : 25 Mar 2019 23:57
Operator : SY/VA
Sample : K2044-14
Misc : 5.38G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BF5W3

Quant Time: Apr 08 10:48:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM031419S.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 26 09:00:24 2019
Response via : Initial Calibration

